



Number 2922

Measuring Inductance Using the 4200-CVU Capacitance-Voltage Unit

Introduction

Although the 4200-CVU capacitance option for Keithley's Model 4200 Semiconductor Characterization System does not measure inductance directly, users can easily extract the inductance from the parameters it does measure: impedance (Z), phase angle (theta or θ), and the test frequency (f). This application note describes how users of the Model 4200-SCS can determine inductance using the 4200-CVU option.

Equipment Requirements

This application requires the use of a Model 4200-SCS system with the 4200-CVU option installed. Although the procedure given here does not depend on the use of a specific KTEI version, 4200-CVU is only supported by KTEI version 7.0 and later.

Implementation procedure

The formulas for deriving the inductance are as follows:

$$X = \frac{Z}{\sin\theta}$$

where:

- X = Reactance
- Z = Impedance
- θ = Phase angle between capacitance and conductance C,G (rads)

$$L = \frac{X}{2\pi f}$$

where:

- L = Inductance
- X = Reactance
- π = Archimedes' constant (3.141592...)
- f = Test frequency

To incorporate these formulas into a KITE project:

- 1. Create an Interactive Test Module (ITM) and choose the 4200-CVU (CVH1 and CVL1) for the measurements at the device terminals in the ITM.
- 2. In the Force Function/Measure Options window, set the Measured Options Parameters to "Z,Theta" as shown in *Figure 1*.
- 3. Open the Formulator and type in the formula for the reactance using the Data Series parameters Z, Theta, and F. The units for "Theta" are in degrees. To use the trigonomic functions in the Formulator, you must convert the units to rads using the "RAD" function as shown in *Figure 2* for a device with terminals A and B:

| Measuring Option: Measured | S | |
|-------------------------------|----------|--|
| Column Names: | | |
| Z_AB | Theta_AB | |
| Parameters: Z,T | V | |

Figure 1



Figure 2

4. Type in the formula for inductance from X and F. "Pi" (π) can be found in the constants area of the Formulator.

Example results

Once the project is saved and executed, the inductance (L) will appear in the Sheet tab as shown in *Figure 3*. It can also be plotted on a graph.

| | A | В | C | D | E | F |
|----|-------------|------------|-----------|------------|-------------|-----------|
| 1 | Z_AB | Theta_AB | DCV_AB | F_AB | X | L |
| 2 | 201.2962E+0 | 89.6496E+0 | 1.0000E+0 | 10.0000E+3 | 201 2925E+0 | 3.2036E-1 |
| 3 | 201.3523E+0 | 99.6733E+0 | 1.0000E+0 | 10.0000E+3 | 201.3491E+0 | 3.2045E-3 |
| 4 | 201.2094E+0 | 89.6547E+0 | 1.0000E+0 | 10.0000E+3 | 201.2057E+0 | 3.2022E-3 |
| 5 | 201.1983E+0 | 89.7035E+0 | 1.0000E+0 | 10.0000E+3 | 201.1956E+0 | 3.2021E-3 |
| 6 | 201.1917E+0 | 89.6774E+0 | 1.0000E+0 | 10.0000E+3 | 201.1885E+0 | 3.2020E-3 |
| 7 | 201.1442E+0 | 89.6390E+0 | 1.0000E+0 | 10.0000E+3 | 201.1402E+0 | 3.2012E-3 |
| 8 | 201.0760E+0 | 89.6673E+0 | 1.0000E+0 | 10.0000E+3 | 201.0726E+0 | 3.2001E-3 |
| 9 | 201.1867E+0 | 89.6774E+0 | 1.0000E+0 | 10.0000E+3 | 201.1835E+0 | 3.2019E-3 |
| 10 | 201.1494E+0 | 89.6973E+0 | 1.0000E+0 | 10.0000E+3 | 201 1456E+0 | 3.2013E-3 |
| 11 | 201 2491E+0 | 89.6900E+0 | 1:0000E+0 | 10:0000E+3 | 201 2452E+0 | 3 2029E-3 |

Figure 3

Other considerations and possible sources of error

This method is useful for measuring inductances ranging from microhenries to millihenries. The actual range that can be measured and the quality of the results achieved will depend on the characteristics of the specific device under test—its own impedance, etc.—and the instrument settings, such as the test frequency. According to the inductance equation, use lower frequencies to measure higher inductances and vice versa.

Standard best practices for high quality electrical measurements should be used, including the use of four-wire connections as close to device as possible, etc.

References

- 1.4200 system overview: http://www.keithley.com/data?asset=50733
- 2.4200 start-up sequence: http://www.keithley.com/data?asset=50696

Specifications are subject to change without notice. All Keithley trademarks and trade names are the property of Keithley Instruments, Inc. All other trademarks and trade names are the property of their respective companies.



A GREATER MEASURE OF CONFIDENCE

KEITHLEY INSTRUMENTS, INC. 🛛 28775 AURORA ROAD 🗖 CLEVELAND, OHIO 44139-1891 🗖 440-248-0400 📮 Fax: 440-248-6168 🔳 1-888-KEITHLEY 🔳 www.keithley.com

BELGIUM

Sint-Pieters-Leeuw Ph: 02-3630040 Fax: 02-3630064 info@keithley.nl www.keithley.nl

ITALY

Peschiera Borromeo (Mi) Ph: 02-5538421 Fax: 02-55384228 info@keithley.it www.keithley.it

CHINA Beijing

Ph: 8610-82255010 Fax: 8610-82255018 china@keithley.com www.keithley.com.cn

JAPAN

Tokyo Ph: 81-3-5733-7555 Fax: 81-3-5733-7556 info.jp@keithley.com www.keithley.jp

SWEDEN

Solna Ph: 08-50904600 Fax: 08-6552610 sweden@keithley.com www.keithley.com

FRAN

Espoo Ph: 09-88171661 Fax: 09-88171662 finland@keithley.com www.keithley.com

KOREA

FINLAND

Seoul Ph: 82-2-574-7778 Fax: 82-2-574-7838 keithley@keithley.co.kr www.keithley.co.kr

SWITZERLAND

Zürich Ph: 044-8219444 Fax: 044-8203081 info@keithley.ch www.keithley.ch

FRANCE

Saint-Aubin Ph: 01-64532020 Fax: 01-60117726 info@keithley.fr www.keithley.fr

MALAYSIA

Penang Ph: 60-4-656-2592 Fax: 60-4-656-3794 chan_patrick@keithley.com www.keithley.com

TAIWAN

Hsinchu Ph: 886-3-572-9077 Fax: 886-3-572-9031 info.kei@keithley.com.tw www.keithley.com.tw

GERMANY

Germering Ph: 089-84930740 Fax: 089-84930734 info@keithley.de www.keithley.de

NETHERLANDS

Gorinchem Ph: 0183-635333 Fax: 0183-630821 info@keithley.nl www.keithley.nl

UNITED KINGDOM

Theale Ph: 0118-9297500 Fax: 0118-9297519 info@keithley.co.uk www.keithley.co.uk

INDIA

Bangalore Ph: 080-26771071,-72,-73 Fax: 080-26771076 support_india@keithley.com www.keithley.com

SINGAPORE

Singapore Ph: 65-6747-9077 Fax: 65-6747-2991 koh_william@keithley.com www.keithley.com.sg

No. 2922